Corrigendum for e-tender reference no. CHEM/2017/400 MHz/01

Title: Supply and Installation of 400 MHz NMR Spectrometer with attachment for solid state and 10 mm BB probes

Details	Present form	Revised form
1. Volume I (page	EMD of appropriate amount	EMD of appropriate amount
no 4)	(Rs.2,50,000/-) on non-	(Rs.4,60,000/-) on non-judicial
	judicial stamp paper of Rs.	stamp paper of Rs. 100/- as per
	100/- as per format in	format in Annexure No. II
	Annexure No. II	
2. Volume I (page	Period of Validity of Bids.	Period of Validity of Bids.
no 8)	Bids shall remain valid for 90	Bids shall remain valid for 150
10 0)	days after the deadline for	days after the deadline for
	submission of bids prescribed	submission of bids prescribed
	by the Purchaser. A bid valid	by the Purchaser. A bid valid
	for a shorter period shall be	for a shorter period shall be
	rejected by the Purchaser as	rejected by the Purchaser as
	nonresponsive.	nonresponsive.
3. Volume I (page	Offer validity Period	Offer validity Period
no 13)	The offer should hold good for	
	a period of <u>90 days</u> from the	The offer should hold good for
	closing date of the tender. Any	a period of <u>150 days</u> from the
	offer falling short of the	closing date of the tender. Any
	validity period is liable for	offer falling short of the validity
	rejection.	period is liable for rejection.
4. Volume I (page	The University of Delhi	The University of Delhi
no 16)	would like to have the	would like to have the
	following time schedule	following time schedule for
	for completion of the	completion of the activities
	activities from the date of	from the date of placement
	placement of orders.	of orders.
	Delivery: 2 months.	Delivery: 6 months.
	Installation,	Installation, commissioning
	commissioning of the	of the equipment, testing &
	equipment, testing &	setting up the unit for
	setting up the unit for	continuous operation must
	continuous operation must	be completed within 2
	be completed within 2 -3	months of the arrival of the
	weeks of the arrival of the	equipment at Dept. of
	equipment at Dept. of	Chemistry, University of
	Chemistry, University of Delhi. It would be	Delhi. It would be
		negotiable, if found
	negotiable, if found	necessary.

	necessary.	
5 Valuma I (naga	The vendor should have a	The vendor should have a
5. Volume I (page no 17)		
	service center in the city of	service center in the city of
	Delhi/Delhi NCR to ensure	Delhi/Delhi NCR to ensure that
	that the machines are attended	the machines are attended
	within a period of 5 hours	within a period of 1 day after
	after the complaint is lodged	the complaint is lodged on
	on working days, and within a	working days, and by the next
	period of 12-24 hours on	working days on holidays.
	holidays. Repairs if any	Repairs if any should be
	should be completed within 48	completed within 48 hours.
	hours.	1
6. Volume I (page	Delay in delivery and	Delay in delivery beyond a
no 18)	installation beyond a period of	period of <u>6 months</u> from the
	<u>6 months</u> from the date of	date of opening of Letter of
	opening of Letter of Credit, or	Credit, or issue of Purchase
	issue of Purchase order	order whichever is later.
	whichever is later.	
7. Volume I (page	During the warranty	During the warranty
no 29)	period of Three years,	period of five years, in
	in case the equipment	case the equipment fails,
	fails, we will provide	the supplier will provide
	all services to complete	all services to complete
	repairs within a week	repairs within a week
8 Volumo I (nago	free of charge.	free of charge.
8. Volume I (page no 32)	Radiofrequency(RF)Generator: Two independent	Radiofrequency(RF)Generator:Twoindependent
110 52)	channels to handle nuclei such	channels to handle nuclei such
	as 1 H, 13 C, 15 N, 19 F, 31 P, etc.	as 1 H, 13 C, 15 N, 19 F, 31 P, etc.
	capable of performing	capable of performing
	multidimensional NMR	multidimensional NMR
	experiments. High	experiments. High performance
	performance power	power transmitters with High
	transmitters with High band	band $({}^{1}\text{H}/{}^{19}\text{F})$ amplifier (100
	$(^{1}\text{H}/^{19}\text{F})$ amplifier (50 watts)	watts) and a low (or Broad)
	and a low (or Broad) band (X)	band (X) amplifier (300 watts
	amplifier (145 watts or more).	or more).

9. Volume I (page no 33)	Solid Sample Accessory with more than 2.5 mm CP/MAS probe. Specify the price for different bore size.	Solid Sample Accessory with more than 2.5 mm CP/MAS probe.
10. Volume I (page no 33)	Spinner for CP/MAS probe (quantity: 10)	Nil
11. Volume I (page no 34)		2 Cryocans of 55 liter capacity with transfer line for N_2 filling.
12. Volume I (page no 34)	20 high quality NMR tubes for solid samples	20 set of rotors with cap and one set of filling tool.