

**Corrigendum to our tender regarding “Specifications for Gel Doc System-UGC/SAP”**

The point # 2 “System must.....excitation.” has been changed to “System must allow for fluorescence based imaging and multiplexing application with the choice of Epi- UV, visible (Red, Green, Blue), preferably with an individual RGB LED’s. Filters to be quoted for multiplexing and should be narrow band pass. System must allow for future upgrade for near IR, IR and deep IR spectral excitation.”

The point #3 “Scientific Grade.....high level of resolution” has been changed to “Scientific Grade CCD camera with sensor size of 1 inch and F0.84 aperture must for high sensitivity and extremely high level of resolution.

The point #12 “4-position filter.... fluorescent stains” has been changed to “6-position Motorized filter wheel allows for dye flexibility of different fluorescent stains”.

The point #13 “Pre-calibrated focus ..... of lens settings” has been changed to “Fixed Motorized optical lens for auto detection of sample height and autofocus”.

The Point #16 “Software should have Apps Studio with library of applications for better ease of use” has been deleted.


In the section titled **Eligibility Criteria** the following two additional points have been added

- Proven after sales service records i.e. company must provide the nearest authorized sales and service representative from the site of installation.
- Company should provide user list with email and Phone numbers of orders placed in the last 2-4years.

In the section titled **Warranty /Comprehensive Maintenance:**

The first point “Machine and accessories .....installation” has been changed to Machine and accessories should be warranted for minimum 2 year from the date of installation certified by the manufacturer.

The second point “The Bidder shall quote for .....place of installation” has been changed to “The Bidder shall quote for under optional item extended warranty upto 3 years after expiry of standard warranty of 2 year. This would cover repair, spare parts etc. supplied by vendor at the place of installation”.



Kaustuv Datta

KAUSTUV DATTA, Ph.D  
Assistant Professor  
Department of Genetics  
University of Delhi South Campus  
New Delhi-110021