

UNIVERSITY COLLEGE OF MEDICAL SCIENCES
& GTB HOSPITAL, SHAHDARA, DELHI – 110095

DEPARTMENT OF ENDOCRINOLOGY
(ICMR Diabetes Cohort Study)

Limited Tender Enquiry

UCMS & GTB H | DEM | R | 2018 | 09

Date: 09.07.2018

Sub: Invitation of bids for supply and installation of Deep freezers (-80⁰ C) (Qty. 2).

Dear Sir/Madam,

Sealed bids are hereby invited for the supply and installation of Two Deep freezers (-80⁰ C) (specifications as per Annexure I). The bid, complete in all respects should reach at the address – Dr SV Madhu, Principal Investigator (P.I.), Head, Department of Endocrinology, Centre for Diabetes Endocrinology & Metabolism, UCMS & GTB Hospital, Dilshad Garden, Delhi – 110095 by 30.07.2018, 4 PM failing which the bid will be treated as late and rejected.

Terms and Conditions

1. Interested bidder should submit their bid in two parts duly sealed in separate envelopes; Part – I, technical bid in one sealed envelope and Part – II, price bid in other sealed envelope. Both the sealed envelopes should then be put in one envelope subscribed as ‘Bid for the supply and installation of Deep freezers (-80⁰ C)’.

Envelope Part- I (Technical Offer): The Technical offer should be complete in all respects and contain all information asked for, except prices. Original catalogue of the model being quoted must also be provided.

Envelope Part - II (Price Bid): Price should be quoted in INR only. TAX should be quoted separately. The Commercial offer must not contradict the technical offer in any way.

2. Earnest Money Deposit: A DD of Rs. 30,000/- (Rupees thirty thousand only) in favour of ‘Principal, University College of Medical Sciences’, payable at Delhi should be submitted along with the technical bid, failing which the bid will be rejected.

In case the successful bidder withdraws his offer within the validity of bid or before entering into the contract within the bid validity period, his EMD will be forfeited. The bidder will have no right to claim for refund of the EMD. The EMD will be returned to the successful bidder after receipt of security money along with the order acceptance from him. The EMD

will be returned to the unsuccessful bidders within one month after placing the order with the successful bidder. No interest will be paid on the EMD amount.

3. Comprehensive onsite warranty for 5 years covering all parts (including consumables) and labour charges from the date of installation. During the warranty period, the bidder will have to undertake comprehensive maintenance of the entire equipment (including components provided by 3rd party, if any) supplied by the bidder. No Traveling allowance or transportation cost will be paid by the purchaser during the warranty period.

The bidder warranty that the goods supplied under this contract are brand new. The entire warranty certificate must be handed over to the P. I.

The bidder shall be fully responsible for the manufacturer's warranty for the equipment, accessories, spare parts etc. against any defects arising from design, material, manufacturing, workmanship, or any act or omission of the manufacturer/ bidder or any defect that may develop under normal use of supplied equipments during the warranty period. This shall also be applicable for any third party tools, software, components which are integrated by the bidder and offered as part of the solution.

Warranty shall not become void even if P.I. decides to install other supplemental component from a third party and installs it with/in these machines.

The bidder shall also be responsible to ensure adequate regular supply of spare parts needed for a specific type of machinery and equipments whether under their annual maintenance and repairs rate contract or otherwise.

Preventive Maintenance: The successful bidder will have to visit for preventive maintenance of the equipment at quarterly intervals during the warranty period.

4. Security Deposit - the successful bidder shall submit 10% (security money) of the cost of the purchase order in the form of DD/PBG issued in favor of '**The Principal, University College of Medical Sciences**' by a scheduled bank after issue of award letter. Only after submission of security money, purchase order shall be issued. The performance bank guarantee shall remain valid till 90 days after the warranty period.

In case the successful bidder fails to provide satisfactory service during the warranty/contract period or found violating bid's terms and conditions, the security money shall not be returned.

Security money shall be released without any interest to the bidder on completion of the bidder's all contractual obligations including the warranty obligations.

5. The equipment should be delivered at the above mentioned address free of cost. No freight charge will be paid.

6. The rates quoted will be binding for a period of one year and supply will be subject to the approval of undersigned.

7. Rates should be quoted in INR only. Applicable Tax should be quoted separately. The payment to successful bidder will be made in INR only after successful installation of the equipment.

8. In case of imported item, successful bidder will be responsible for custom clearance of the equipment. Also, no extra charges would be borne by the institute towards these expenses.

9. Undersigned reserves the right to accept the quotations in part / full or reject them even if they are lowest.

Thank You,



(Dr. S V Madhu)
P. I.

Specifications for Deep Freezer (-80⁰ C)

1. Cabinet of the equipment should be upright.
2. Equipment should be microprocessor controlled.
3. Adjustable temperature range of upto – 80⁰ C or low.
4. Capacity of the equipment should be 450 ±50 ltr.
5. Interior of the equipment should be of stainless steel.
6. The equipment should be equipped with VIP insulation panels.
7. There should be a minimum of four shelves/chambers in the equipment.
8. Equipment should be capable of providing desired temperature of up to –80⁰ C at ambient temperature of ≥30⁰ C.
9. Audible and visual alarm system to indicate high temperature and power failure.
10. Refrigerant should be CFC free.
11. Equipment should have castors for easy mobility.
12. Battery backup for monitoring of temperature in case of power failure.
13. External dimension should not be more than 950 mm * 950 mm * 2000 mm (W*D*H).
14. Equipment should be US FDA approved and or European CE certified.
15. Equipment should be provided with compatible CVT.
16. Equipment should be able to operate at power supply: 220-240 V.

