

UNIVERSITY OF DELHI SOUTH CAMPUS
Centre for Innovation in Infectious Disease Research, Education and Training (CIIDRET)

First Floor Engineering Department
New Delhi-110021
Tel.: 91-11-24115883
E-mail: vkchaudhary@south.du.ac.in

E-tender Notice

Ref. No.: UDSC/CIIDRET/VKC/2018/004

July 23, 2018

Tenders are hereby invited in two-bid system from manufactures or their authorized dealers for the **supply of Chest Deep Freezer [Quantity: Four (04) numbers]** with the specification given below:

1. The Deep Freezer should have 2 insulated doors with independent locking system.
2. The Deep Freezer should have digital display and temperature controller with features to set differential.
3. The temperature should go at least up to -20°C.
4. The Net capacity should be ~ 380 liters.
5. Refrigerant: R-134 A
6. External Dimensions in inches (Width: 57.5, Depth: 26.5, Height: 33)
7. Internal Dimensions in inches (Width: 52.2, Depth: 19.9, usable Height in main section: 25.5)
8. Inner cabinet should be made of Pre-Painted steel
9. Wheels and Door Lockable option
10. Power Supply: 220V to 240V - 50HZ
11. Power Consumption: 210 Watts
12. PUF insulation of 60 mm thickness
13. The compressor should carry 3-years warranty
14. Should come with one basket.

Terms and Conditions:

1. The bids should be uploaded on to the e-procurement website (<https://eprocure.gov.in>) within 10 days of the date of the advertisement (latest by 2nd August 2018). The Bidders/Vendors are advised to follow the instructions provided in the 'Instructions for Online Bid Submission' for the e-submission of the bids online through the Central Public Procurement Portal for e-Procurement at <https://eprocure.gov.in/eprocure/app>.
2. Quotations have to be submitted in two bid systems. The First part 'Technical bid', should consists of all technical details and supporting documents with terms and conditions. A compliance sheet (in the attached format; Annexure I) must be filled by the vendor against each point and giving reference of the same (page no., line no.) in the supporting company brochure/document. **Also mention the Model and make of the machine.**
3. The second part 'Financial bid', should contain item-wise pricing of items mentioned in the technical bid. The Financial Quotations should contain price of the equipment, discount if any, packaging and forwarding charges, Air Freight and insurance charges. The price quoted should be F.O.R destination price in INR against customs duty exemption certificates provided by the buyer and inclusive of standard installation.
4. **Manual bids shall not be accepted. Do not submit manual bids by hand.**
5. It may be noted that the Institution is eligible for concessional as per the Notification No. 45/2017 and 47/2017-Central Tax (Rate) Dt. 14th November, 2017, issued by the ministry of Finance Department of Revenue of Govt. of India will be acceptable, if applicable on the items in the tender. The price quote should keep this in the view. Therefore, quote price without GST, and the GST will be paid as applicable. If we would pay concessional GST, appropriate certificate will be provided.
6. The supplier should also provide a list of at least five users (name, phone number and email addresses) in academic Institutions for the item quote in the bid. Also, provide certificate from the existing users (at least three) about the performance of the machine, quality of services and other support.
7. Payment will be made by RTGS mode within 30 days of satisfactory installation of the Freezer. The supplier/OEM will be responsible for the installation and providing service support during the warranty period.
8. The quote should be valid for 90 days from the due date.
9. Authorization certificate from the manufacturer should be attached.

निदेशक (CIIDRET)
नवपरिवर्तन केंद्र - संशोधन और प्रशिक्षण
Center for Innovation in Infectious Disease
Research, Education & Training (CIIDRET)
दिल्ली विश्वविद्यालय दक्षिण परिसर, नई दिल्ली - 110021
University of Delhi South Campus, New Delhi-110021

10. The institute reserves the right to order equipment with better quality and suitability over lower price and to accept or reject any or all quotations without assigning reasons thereof. The tender may also be cancelled without providing reasons.
11. The tenderers are requested not to attach any additional paper more than requested in the terms and conditions.
12. **The tenderers are requested not to attach any additional paper more than requested in the terms and conditions.**
13. **Tenderer should attach in the technical bid, a print out of detailed price bid with original part no, quantities but without mentioning the quoted price. Technical bid without inclusion of this document will be considered incomplete**
14. **EMD of Rs. 2000 (two thousands only) payable to "Director, UDSC" at New Delhi, in the form of bank draft /bankers cheque in a sealed cover with tender detail and number mentioned should reach Professor Vijay K Chaudhary, Director, CIIDRET, University of Delhi South campus, First Floor Engineering Department Building, New Delhi-110021 before the end date and time of bid submission. The tender without EMD will not be considered.** Bidder, however have to attach scanned copies of EMD proof along with their e-tender.
15. For any query, please contact Professor Vijay K Chaudhary, Director, CIIDRET, University of Delhi South campus at 24115863/vkchaudhary@south.du.ac.in.


Professor Vijay K Chaudhary
Director, CIIDRET

निदेशक / Director
नवपरिवर्तन केन्द्र -संक्रामक रोग अनुसंधान, शिक्षा और प्रशिक्षण
Center for Innovation in Infectious Disease
Research, Education & Training (CIIDRET)
दिल्ली विश्वविद्यालय दक्षिण परिसर, नई दिल्ली - 110021
University of Delhi South Campus, New Delhi-110021

Annexure I

Sheet for technical compliance for Chest type Deep Freezer

S. No	Feature	% Compliance (describe deviation)	Reference of the specification (page no., line no. in the product catalogue)
1.	The Deep Freezer should have 2 insulated doors with independent locking system.		
2.	The Deep Freezer should have digital display and temperature controller with features to set differential.		
3.	The temperature should go up to -25°C.		
4.	The Net capacity should be ~ 380 litres.		
5.	Refrigerant: R-134 A		
6.	External Dimensions (Width: 57.5, Depth: 26.5, Height: 33) in inches		
7.	Internal Dimensions (Width: 52.2, Depth: 19.9, usable Height in main section: 25.5) in inches		
8.	Inner cabinet should be made of Pre-Painted steel		
9.	Wheels and Door Lockable option		
10.	Power Supply: 220V to 240V - 50HZ		
11.	Power Consumption: 210 Watts		
12.	PUF insulation of 60 mm thickness		
13.	The compressor should carry 3-years warranty		
14.	Should come with one basket.		
15.	Mention Model No.		
16.	Tenderer should attach in the technical bid, a print out of detailed price bid with original part no, quantities but <u>without mentioning the quoted price</u>. Technical bid without inclusion of this document will be considered incomplete.		

Additional eligibility criteria

Manufacturer should have a nationwide dealer network with strong technical/Repair support available locally and provide documentary evidence for the same.




निदेशक / Director
 नवपरिवर्तन केन्द्र - संक्रामक रोग अनुसंधान, शिक्षा और प्रशिक्षण
 Center for Innovation in Infectious Disease
 Research, Education & Training (CIIDRET)
 दिल्ली विश्वविद्यालय दक्षिण परिसर, नई दिल्ली - 110029
 University of Delhi South Campus, New Delhi-110021

Annexure II

Format for Price Bid for the supply of Chest Deep Freezer [Quantity: Four (04) numbers]

S. No.	Item (include part no)	Price (Rs./ foreign currency)*
1.	Cost of 4 nos. of Chest Deep Freezer (Model No.....)	
	Total (without Tax)/CIF (Air)	
	Tax (rate)	
	Total with Tax	

*All cost should be without added Taxes. It may be noted that the University of Delhi is eligible for both concessional GST (@5%) as per the Notification No. 45/2017 and 47/2017-Central Tax (Rate) Dt. 14th November, 2017, issued by the ministry of Finance Department of Revenue of Govt. of India and concessional Custom duty. However, the applicable taxes/duties should be mentioned at the current prevailing rates keeping in view the current exemptions for the University of Delhi. For direct import, University will get the custom clearance done.



निदेशक / Director
नवपरिवर्तन केन्द्र - संक्रामक रोग अनुसंधान, शिक्षा और प्रशिक्षण
Center for Innovation in Infectious Disease
Research, Education & Training (CIIDRET)
दिल्ली विश्वविद्यालय दक्षिण परिसर, नई दिल्ली - 110029
University of Delhi South Campus, New Delhi-110021