

SPECIFICATION FOR REFRIGERATED CIRCULATORY WATER BATH

The following features must be standard part of the system and documentary evidence of the same should be provided along with technical bid.

- Abrasion Resistance and durable finish
- Water Bath Capacity: 6-8 liters
- Temperature range: -10°C to 100°C
- Control Accuracy : $\pm 0.1^{\circ}\text{C}$
- Power Supply : 230 VAC, 50 Hz, Single Phase
- Should have provision to connect to external instruments to maintain temperature. All accessories, parts, connection pipes, nozzles etc for connecting to other instrument should be a part of supply
- Uniformity : Better than $\pm 0.3^{\circ}\text{C}$
- Digital Temperature Controller Unit
- Product should be ISO , CE certified
- System should be supplied with basic spares

Eligibility Criteria

- Bidder should be Manufacturer/ Authorized Partner/ Reseller of the manufacturer and a Letter of Authorization from manufacturer for the same and specific to the tender should also be enclosed. The bidder should also be the Authorized Service Provider.
- Bidder should be financially sound to execute the order. Certificate to this effect should be issued by any nationalized/ scheduled bank showing its annual turnover of at least 25Lakhs each in the last three financial years.
- The Bidder shall provide the Registration number of the firm along with the LST/ CST/ WCT No. and the PAN Number issued by the concerned authorities.
- Separate technical and price bid should be uploaded.
- Technical Bid must contain a Technical Compliance sheet. Technical compliance sheet must have the page number of a published catalogue/operating manual/manufacture website as proof for compliance with the instrument and accessories for each technical points. Relevant copies of catalogue/operating manual/manufacture website pages that have been listed as proof should be attached.
- Price Bids should contain the prices, terms of delivery, sales, payment terms etc.
- The bidder must not be blacklisted by Delhi University. A Certificate or undertaking to this effect must be submitted.

No Commitment to Accept Lowest or Any Tender

- Demonstration of the equipment will be required if needed.
- University of Delhi shall be under no obligation to accept the lowest or any other offer received in response to this tender notice and shall be entitled to reject any or all offers. University of Delhi will not be obliged to meet and have discussions with any vendor, and or to listen to any representations.

Warranty /Comprehensive Maintenance

- Machine should be warranted for minimum of 12 months from the date of installation.
- Installation and commissioning to be completed up to the satisfaction level of the technical purchase committee followed by training and handling by the bidders

Price quotation

- F.O.R. UDSC price (INR/FC) to be quoted which includes, custom duty & clearance as well as transportation charges to site of installation. CDEC will be provided.
- Payment will be made against successful installation.
- Validity of the quotation should be for 6 months.


Published Date: 22-12-2018 (14:00 Hrs)

Bid Document Download Start Date: 24-12-2018 (10:00 Hrs)

Bid Submission Start Date: 25-12-2018 (10:00 Hrs)

Bid Submission End Date: 16-01-2019 (17:00 Hrs)

Bid Opening Date: 18-01-2019 (11:00 Hrs)



Kaustuv Datta

KAUSTUV DATTA, Ph.D
Assistant Professor
Department of Genetics
University of Delhi South Campus
New Delhi-110021

All the relevant documents related to tender can be found on CPPP portal