

DEPARTMENT OF PLANT MOLECULAR BIOLOGY
UNIVERSITY OF DELHI SOUTH CAMPUS
NEW DELHI – 110 021

Subject: Notice inviting "Quotations for supply of Contamination Monitor CM-10"

Please quote your lowest rates for **supply of Contamination Monitor CM-10 – One** with following Specifications:

Sensor	Pancake GM tube based detector
Probe Configuration	The detector is in an external cylindrical aluminium housing, connected to the main unit via a 1m cable. A removable cover is provided for beta detection
Radiation Detected	α , β and γ
Sensitivity	60 CPS/mR/hr
Measurement Units	CPS, CPM, mR/hr, uSv/hr
Measuring Ranges	0-6000 CPS
	0-360000 CPM
	0.001-100 mR/hr
	0.01-1000 μ Sv/hr
Accuracy	+/- 15% with Cs ¹³⁷
Resolution	1 CPS, 1 CPM, 0.001 mR/hr, 0.01 uSv/hr
Alarm	<ul style="list-style-type: none"> - User selectable alarm threshold - Beeping audio (with counts) when reading < threshold - Continuous audi when rading > threshold - Blinking Backlight when reading > threshold
Display	High resolution graphic lcd (45.2x27 mm) with white backlight
Keypad	Tactile swithches with a sealed membrane
Power	4 x 1.5 V AA Alkaline batteries
Battery Life	>60 hours
Device Enclosure	Water resistant plastic enclosure
Operating Temperature	0°C to 50°C
Size	80 x 125 x 36 mm
Weight	500 g with sensor
Warranty	One year

<<ORGANISATION NAME>>

FINANCIAL BID

S.No	Description	Qty	Rate	Total Amount (Rs) (inclusive of GST 5%)
1	CONTAMINATION MONITOR CM-10 (Specification as per tender notice)	1 no.		

Conditions:-

1. The University authorities reserve the right to reject any or all quotation without assigning any reason.
2. Note: “ 5% GST is applicable for University of Delhi as per the DSIR registration No. TU/V/RG-CDE (14)/2016 dated 16-03-2018, which is in accordance with Govt. Notification No. 47/2017-Integrated tax (Rate) dt. 14-11-2017, and Notfn. No. 45/2017-Central Tax (Rate) dt. 14.11.2017, and Notfn. No. 45/2017-Union Territory Tax (Rate) dt.14-11-2017.”

Eligibility Criteria

- Bidder should be Manufacturer/ Authorized Partner/ Reseller of the manufacturer and a Letter of Authorization from manufacturer for the same and specific to the tender should also be enclosed. The bidder should also be the Authorized Service Provider.
- Bidder should be financially sound to execute the order.
- **The Bidder shall provide the Registration number of the firm along with the LST/CST/ WCT No. and the PAN Number issued by the concerned authorities.**
- **The bidder should enclose similar orders provided to various reputed Government Institutes/Research labs/ICAR/CSIR institutes throughout India in the last 5 years.**
- Separate technical and price bid should be uploaded.
- Technical Bid must contain a Technical Compliance sheet. Technical compliance sheet must have the page number of a published catalogue/operating manual/ manufacture website as proof for compliance with the instrument and accessories for each technical points. Relevant copies of catalogue/operating manual/ manufacture website pages that have been listed as proof should be attached.


- All vendors are requested to attach original technical literature/ catalogue in support of the mentioned specifications & highlight the above features. The same features should be available in the company website.
- Price Bids should contain the prices, terms of delivery, sales, payment terms etc.
- The bidder must not be blacklisted by Delhi University. A Certificate or undertaking to this effect must be submitted.
- Payment will be made by RTGS after installation of instrument.

No Commitment to Accept Lowest or Any Tender

- Demonstration of the equipment will be required if needed.
- University of Delhi shall be under no obligation to accept the lowest or any other offer received in response to this tender notice and shall be entitled to reject any or all offers. University of Delhi will not be obliged to meet and have discussions with any vendor, and or to listen to any representations.

Warranty /Comprehensive Maintenance

- Machine should be warranted for minimum 12 months from the date of installation.
- Additional 1 year warranty from date of installation should be quoted as an optional item in the price bid.
- Installation and commissioning to be completed up to the satisfaction level of the technical purchase committee followed by training and handling by the bidders Price quotation
- F.O.R. UDSC price only in INR to be quoted which includes transportation charges to site of installation. IGST/CDEC will be provided.
- Payment will be made against successful installation.



HEAD OF THE DEPARTMENT