

DEPARTMENT OF PLANT MOLECULAR BIOLOGY
UNIVERSITY OF DELHI SOUTH CAMPUS
NEW DELHI – 110 021

Ref. No.PMB/CARI/F/15-16

February 2, 2016

M/s _____

Dear Sirs:

Kindly quote (both technical and financial bid separately) your lowest rates for the supply of "Incubator Shaker" (qty: 02 nos) with following specification:

1. Incubator shaker with microprocessor digital PID control for precise programmable temperature and speed.
2. Sensitive membrane type keypad with timer and LED 4 digital display.
3. Temperature range 10°C to 60°C with set accuracy $\pm 0.1^\circ\text{C}$ with uniformity 0.5°C and rpm range 0 to 300 or more with 20 mm stroke orbital motion.
4. Electronic controlled shaking mechanism for quiet orbital motion, precise speed control and shaking motion automatically stopping when door is opened.
5. Feature like over- temperature cut-off and over-current breaker.
6. Universal spring rack/different capacity clamps accommodating various sized flasks up to 2L and 50ml and 15ml tube racks easy to attach and remove from the shaking plate.
7. In case of disruption in power supply, shaker should continue shaking automatically on resumption of power supply.
8. Shaker operational at power supply of 220 V / 50-60 Hz.
9. Minimum two year warranty with provision of extension

Please Note:

- *Technical bid must include compliance sheet, list of previous installations (with contact details) and related published reference documents/web links in support of your claims.*
- *Offer with incomplete information will be rejected.*
- *100% payment would be made after delivery and satisfactory installation of the product.*

Your sealed offer should reach the undersigned latest by February 21, 2016.

Thanking you,

Yours sincerely

Jitendra Khurana

Programme Co-ordinator

2.2.2016

Jitendra Khurana
2.2.2016

Professor J.P. Khurana
Head
Department of Plant Molecular Biology
University of Delhi South Campus
New Delhi-110021

Am.